



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105448	PRINTER SN 09B.3591.016	DATE OF INSPECTION 02/04/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Holden Police Department 110 West 3rd street Holden, Mo 64040		TIME OF INSPECTION 9:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☒ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER Guth Labs LOT # 13210 EXP. DATE 07/29/2015

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2252 SIMULATOR EXP DATE _____

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098

TEST 2 .098

TEST 3 .099

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Chief Richard Martin
TYPE II PERMIT NUMBER/EXPIRATION DATE 220428 12/27/2015	TELEPHONE NUMBER (816) 850-4154

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 185448
Version no: 532B

TEST RECORD 00165

Temp Date Time ^{s/} 210L

Air Blank:
02/04/14 09:35 .000
Calibration Check:
26 02/04/14 09:35 .098

Subject Name

MAINTENANCE TEST

Subject I.D.

TEST #1

Operator Name, I.D.

MARTIN 220428

Location

Howe PD

AS IV Serial no: 185448
Version no: 532B

TEST RECORD 00166

Temp Date Time ^{s/} 210L

Air Blank:
02/04/14 09:36 .000
Calibration Check:
27 02/04/14 09:36 .098

Subject Name

MAINTENANCE TEST

Subject I.D.

TEST #2

Operator Name, I.D.

MARTIN 220428

Location

Howe PD

AS IV Serial no: 185448
Version no: 532B

TEST RECORD 00167

Temp Date Time ^{s/} 210L

Air Blank:
02/04/14 09:39 .000
Calibration Check:
28 02/04/14 09:39 .099

Subject Name

MAINTENANCE TEST

Subject I.D.

TEST #3

Operator Name, I.D.

MARTIN 220428

Location

Howe PD

AS IV Serial no: 185448
Version no: 532B

TEST RECORD 00168

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/04/14 09:40

Subject Name

MAINTENANCE TEST

Subject I.D.

RFI TEST

Operator Name, I.D.

MARTIN 220428

Location

Howe PD